

EC TYPE-EXAMINATION CERTIFICATE: 2031

Product description: - Filtering Facepiece Particulate Respirators.

Product identification: - 3M 8000 Series - See Schedule.


Manufacturer: - 3M United Kingdom PLC
3M Centre
Bracknell
Berkshire
RG12 8HT
United Kingdom

When assessed and examined against harmonised standard EN149:2001 incorporating amendment no. A1:2009 are found to be in conformity with Council Directive 89/686/EEC and associated amendments, relating to personal protective equipment.

And

Are manufactured under a Quality Control System which has been satisfactorily assessed as meeting the requirements of Article 11 Section B of the same Directive.

Signed



Date: 29th October 2015


K J Warren, Manager, Certification Services

For and on behalf of INSPEC International Ltd.

56 Leslie Hough Way, Salford, Gt Manchester M6 6AJ, England (Notified Body No: 0194)

Certification remains valid until 26th September 2018, providing compliance with INSPEC's terms and conditions is maintained.

Original date of issue: 16th March 2010



For terms and conditions of issue, see page 2

Terms and Conditions

Reference Documents: -

- | | | | |
|------|--------------------------|---|---|
| i) | Test Reports | - | 3M Test reports 09212, 09213, 09214, 09215, 09061, 09211, 14044, 15151B and 15152B. |
| ii) | Technical File | - | TF/2031.
3M Tech File TF0383 refers. |
| iii) | Test and Inspection Plan | - | RSF-8710, RSF-8810, RSF-8812, RSF-8820, RSF-8822 and RSF-8832. |

Conditions attached to the issue of this certificate:

- i) Marking and instructions have been assessed in the English language only. It is the Manufacturers/Authorised Representatives responsibility to obtain and supply language versions acceptable to the country where the product is to be sold.
- ii) Any changes to the product, technical file or quality manual/quality plan shall be immediately notified to INSPEC.
- iii) The Manufacturer/Authorised Representative shall comply at all times with INSPEC's Regulations governing CE Product Certification.
- iv) Satisfactory maintenance of independent certification against Article 11.B of the P.P.E. Directive.
- v) This Certificate remains the property of INSPEC and may be withdrawn if any of the conditions attached to its issue are not complied with.
- vi) This certificate may be copied or reproduced by the certificate holder, complete and without omissions or additions, and in accordance with INSPEC's terms of business.
- vii) This certificate has been provided in accordance with our standard Terms of Business, which can be viewed at, and printed from:

<http://inspec-international.com/ToB.pdf>

If you have difficulty accessing the Terms of Business, you may contact us for a copy.

CERTIFICATION INDEX

Item	Status	Issued	Amendment
Pages 1 – 4	Withdrawn	100316	Initial issue
Pages 1 – 4	Withdrawn	120314	Addition of 3M Aycliffe Report and change to release specification. Amendment to Terms and Conditions.
Pages 1 – 4	Withdrawn	130926	This certificate was originally for 8710E/06983 and 8710 models only. This amendment includes models 8810/06988, 8812, 8820 and 8822/06922 now merged onto this certificate from certificate numbers 2032, 2033, 2034 and 2035, with the models listed now recorded in a single 3M Technical File (TF0383). This amendment also addresses a change in preform material, reverting to previous product construction across the models listed. Addition of five year validity rule being applied to Certificate.
Pages 1 – 4	Withdrawn	140721	Addition of model 8832, RSF-8832 and 3M Aycliffe test report 14044.
Pages 1 – 4	Valid	151029	Addition of 3M Aycliffe test reports 15151B and 15152B.

CERTIFICATION SCHEDULE

Filtering Facepiece Particulate Respirator

Model	Class	Valved/Non-Valved
-------	-------	-------------------

8710E	FFP1 NR D	Non-valved
-------	-----------	------------

The above model is also supplied as model reference 06983

8710S	FFP1 NR D	Non-valved
-------	-----------	------------

8810	FFP2 NR D	Non-valved
------	-----------	------------

The above model is also supplied as model reference 06988

8812	FFP1 NR D	Valved
------	-----------	--------

8820	FFP2 NR D	Non-valved
------	-----------	------------

8822	FFP2 NR D	Valved
------	-----------	--------

The above model is also supplied as model reference 06922

8832	FFP3 NR D	Valved
------	-----------	--------

In addition to the products referenced on this certificate, the standard product may also be sold as market specific packaged variants.

Such variants will be differentiated and the Technical File will be updated with the appropriate information.

